

LAUREL[®] ELECTRONICS, LLC
MEASUREMENT & CONTROL INSTRUMENTS
3183-G Airway Avenue, Costa Mesa, CA 92626, USA

• Tel: (714) 434-6131 • Fax: (714) 434-3766 • Website: www.laurels.com • Email: sales@laurels.com

Certificate of Calibration

Laurel Electronics, LLC
Certifies that this unit

Model L7041AFR Serial Numbers 15115

has had all ranges tested and calibrated to meet or exceed published specifications for accuracy and performance. The calibration was performed with one or more of the following traceable to NIST standards.

Calibration correction stored in meter nonvolatile memory -43 PPM

Equipment used:

Model	Serial Number	Calibrated	Due	NIST Traceability
<u>HP10544A</u>	<u>8961641</u>	<u>21Feb25</u>	<u>21Feb26</u>	<u>134911</u>

Environmental:

Temperature: 23 to 27°C

Humidity: < 95% RH

Date 25-November-2025

Signed _____